



World Class Accreditation

The American Association for Laboratory Accreditation

Accredited Laboratory

A2LA has accredited

DICKSON

Addison, IL

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).

Presented this 17th day of June 2011.





Peter Meyer

President & CEO
For the Accreditation Council
Certificate Number 1621.01
Valid to July 31, 2013

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

DICKSON
 930 South Westwood Ave.
 Addison, IL 60101
 Karen Dargenio Phone: 630 543 3747

CALIBRATION

Valid To: July 31, 2013

Certificate Number: 1621.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Thermodynamics

Parameter/Equipment	Range	CMC ² (±)	Comments
Humidity	(0.2 to 89.99) % RH at (15 to 35) °C	0.6 % RH	Chilled mirror standards
	(90 to 99.5) % RH at (15 to 35) °C	0.7 % RH	Relative humidity (RH)
Resistance Thermometry	(-60 to 400) °F	0.2 °F	Platinum RTD
Thermocouple Temperature Simulation – Type K	(-200 to 2000) °F	0.9 °F	Thermocouple simulator

¹ This laboratory offers commercial calibration service for Dickson products only.

² Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Peter Meyer
 Page 1 of 1